

CONVENTUS DRS INTRAMEDULLARY FIXATION FOR DISTAL RADIUS FRACTURES: A CASE SERIES

Stephanie Malliaris MD, Schneider K. Rancy BA, Scott W. Wolfe, MD
Hospital for Special Surgery, New York, NY USA



Introduction

Conventus DRS is a novel FDA-approved intramedullary fracture fixation device used for distal radius fractures. Volar plating complication rates range 3-37% in the literature. Conventus DRS's complication rate is reported as 8.3% in the company's white paper. We examine a series of 20 patients with distal radius fractures treated with Conventus DRS ORIF and report the outcomes.

Methods

20 patients with distal radius fractures (15 AO type C1 and C2; 5 AO type A2 and A3) underwent ORIF using Conventus DRS by a single surgeon. Data including fracture type, angle of displacement, grip strength, wrist ROM, PRWE, and VAS pain scores were recorded at time of surgery and at follow-up.

*Average age: 48 yrs (18-74)

*Follow up: 4.6 mos (2-12)

Results



Figure 1. Postoperative PA and lateral radiographs.

Table 1. Objective Results.

Parameter	Affected	Contralateral
Extension (deg)	76 (91%)	83
Flexion (deg)	73 (86%)	86
Rad Dev (deg)	15 (86%)	17
Ulnar Dev (deg)	29 (86%)	34
Supination (deg)	79 (98%)	80
Pronation (deg)	77 (98%)	78
Grip Strength (lbs)	54 (75%)	76
Digital Flexion	0 (100%)	0

Table 2. Radiographic Results.

Parameter	PreOp	2 Wks	Final
Ulnar Variance (mm)	2.2	-0.5	-0.3
Radial Inclination (deg)	22	21	22
Volar Tilt (deg)	-24	-2	2
Coronal Shift (mm)	4.0	0.7	0.1
DRS Axle to Subchondral Plate (mm)	N/A	1.7	1.6

- VAS Pain Score: average = 0.8 (median 0)
- PRWE: average = 19 (range 0-64)
- No tourniquet was used
- All patients: full digital motion at first follow-up visit

Complications

Screw Removal – 2 pts
 1 - asymptomatic
 1 - dorsal tendon irritation
 Hardware Removal – 1 pt
 fracture subsidence
 Screw Prominence – 1 pt
 Dorsal Tendon Irritation (screw removal pending)
 No infections
 No CRPS

Conclusions

Conventus DRS is a promising alternative to established methods of distal radius ORIF.

- Minimal postoperative swelling
- Rapid return of digital function
- Low pain scores
- Lack of soft tissue complications
- Safe and effective for articular and extra-articular distal radius fractures
- Learning curve for device